

#3

# ABSTRACT

Electrospray mass spectrometry has been shown to be useful and extremely accurate (about 1 mass unit/10,000 MW) to a total mass of 30-40 kD. Data clearly shows that use of electrospray mass spectrometry and protein sequencing as applied to the identification of ETEC CF.

2011-07-27 14:50